

Flame Retardant Declaration of Conformity

Manufacturer:
 Jiangyin Everise Medical Equipment Co.,Ltd
 1001 Chengyang Road,Jiangyin City,
 Jiangsu,China,

Representative:
 Zhongtian Textile Testing Co.,Ltd
 No. 18, Jingdu 2nd Road, Warp Knitting Industrial
 Park,Haining City,Zhejiang,China

We, the manufacturer, herewith declare that the products:

Folding Stretcher fabric


Test Report

Report No. :W2018-08452

Test Result:

Test Project			Test Method	Technical Requirement	Test Result	Single Verdict	Remark
Combustion performance	Damage length /mm	Radial direction	GB/T 5455-2014	≤200	58	Confirm	The sample when burned, no dripping objects
		Zonal direction		≤200	64	Confirm	
	Keeping burning time /s	Radial direction		≤15	0	Confirm	
		Zonal direction		≤15	3	Confirm	
	Flame retardant time /s	Radial direction		≤15	8.8	Confirm	
		Zonal direction		≤15	7.6	Confirm	

Sample photo



W2018-08452

*****End of report*****

We herewith declare that the above mentioned products meet the fireproof and flame retardant standard. All supporting documentations are retained under the premises of the manufacturer. We are exclusively responsible for the Declaration of Conformity.

The above mentioned declaration of conformity is exclusively under the responsibility of

Jiangyin Everise Medical Equipment Co.,Ltd
 1001 Chengyang Road,Jiangyin City,Jiangsu,China

Dongjiang G
 Director of Quality
 October 2023

江阴万事兴医疗器械股份有限公司
 JIANGYIN EVERISE MEDICAL EQUIPMENT CO.,LTD