

EC DECLARATION OF CONFORMITY	
Product Group	Absorbable Surgical Suture (Synthetic)
Model/Type	Absorbable Surgical Suture
GMDN Code	GMDN Code: 45814
Brand Name	Monoglyde
Generic Name	Absorbable Surgical Suture USP (Synthetic) Monofilament Poliglecaprone 25
Sizes: 8-0, 7-0, 6-0, 5-0, 4-0, 3-0, 2-0, 0, 1, 2	
Indented Use: Monoglyde sutures are intended for use in general soft tissue approximation, but not for use in cardiovascular or neurological tissues, microsurgery or ophthalmic surgery.	
Quality Responsibility: Healthium Medtech Pvt Ltd., Bangalore declares under Sole responsibility for complying all The quality requirements of the product.	
Address (office and Factory) of the Manufacturer:	Plot No.1600, R – 6 West, Sri City (SEZ), Cherivi Village, Irrugulam Post, Satyavedu Mandal, Chittoor Dist, Andhra Pradesh – 517588, India Ph: +91-80-41868000 – 41868030
E.U. Representative's name and address	MED DEVICES LIFESCIENCES B.V. Kraijenhoffstraat 137 A, 1018RG Amsterdam, Netherlands Email: info@meddevices.net Phone: +31-202254558
Classification of the device:	Monoglyde is classified as Class III device in accordance with Rule 8 of Annex IX of the MDD 93/42/EEC.
Conformity Assessment Route	Annexure II, excluding (4) (Module H) of MDD 93/42/EEC Council directives as amended by 2007/47/EC
Notified Body's Address and	3EC International a.s. Hranicna 18, 821 05 Bratislava, Slovak Republic.
Notified body number	NB 2265
Certificate No.:	2017-MDD/QS-055/B
Document Reference No.	TF/01/D
Applicable Standards	EN ISO 13485:2016, EN ISO 14971:2012, EN ISO 15223-1 : 2016 , EN 1041:2008, EN 556-1: 2001, EN 556-2: 2015, ISO 11135: 2014, ISO 10993-1:2009, ISO 10993-5: 2009, EN ISO-10993-6 :2009, ISO 10993-7: 2008, ISO 10993-10: 2010, ISO 10993-13: 2010, EN ISO-10993-11 : 2009, EN ISO-11607-1 : 2009, EN ISO-11607-2 : 2006, EN ISO-11737-1: 2006, EN ISO 11737-2: 2009, EN ISO-14630 : 2012, ISO 14644-1:2015, EN 868-7:2017, MEDDEV 2.7-1 rev 4, MEDDEV 2.12-1 rev 8, MEDDEV 2.12-2 rev 2
Declaration: Healthium Medtech Pvt Ltd. Declares that Monoglyde meets the requirements of Quality as per USP specifications given in the monograph, Absorbable Surgical Suture, Synthetic and also complies with the provisions of the council of Directives 93/42/EEC as amended by 2007/47/EC Directives for Medical devices.	
Signature :	
Name and Position	Malesh. M Regulatory Affairs
Date	27.03.2020
Place	Bangalore